



LABORATORY TESTING INC.

2331 Topaz Drive, Hatfield, PA 19440
TEL: 800-219-9095 • FAX: 800-219-9096

Certified Test Report

EUL001-13-06-20939-1



SOLD TO

Echo Ultrasonics, LLC
2457 West Shore Drive
Lummi Island, WA 98262

SHIP TO

Echo Ultrasonics, LLC
2457 West Shore Drive
Lummi Island, WA 98262
ATTN: Gene Larson

CUSTOMER P.O.

13331

CERTIFICATION DATE

6/18/2013

SHIP VIA

EMAIL

DESCRIPTION

Quantity: 1
Description: VersaSonic Ultrasonic Couplant
Batch: 801

CHEMICAL ANALYSIS:

APPLICABLE SPECIFICATIONS: Customer Instructions

KEY: C - Conforms NC - Non-Conformance R-Report for Information

<u>ELEMENT</u>	<u>ACTUAL</u>
Br + Cl + I	48 ppm
F	<1 ppm
S	15 ppm

KEY (C/NC/R): R

Procedures/Methods: MAS-ICP, Rev. 10, Analysis of Metals & Their Alloys by ICP; 86-WC-1, Rev. 3, Analysis by Classical Wet Chemistry

The services performed above were done in accordance with LTI's Quality System Program Manual Revision 20 dated 12/12/12 and ISO/IEC 17025. These results relate only to the items tested and this report shall not be reproduced, except in full, without the written approval of Laboratory Testing, Inc. L.T.I. is accredited by Nadcap for NDT and Materials Testing for the test methods and specific services as listed in the Scopes of Accreditation available at www.labtesting.com and www.eAuditNet.com. The results reported on this test report represent the actual attributes of the material tested and indicate full compliance with all applicable specification and contract requirements.

MERCURY CONTAMINATION: During the testing and inspection, the product did not come in direct contact with mercury or any of its compounds nor with any mercury containing devices employing a single boundary of containment.

NOTE: The recording of false, fictitious or fraudulent statements or entries on this document may be punishable as a felony under Federal Statutes.

Cindy Heckler
QA Supervisor
Cindy Heckler
Authorized Signature ^(ES)